

统一社会信用代码

914406081935502519

# 营业执照

(副本) (副本号:2-1)



扫描二维码登录“国家企业信用信息公示系统”了解更多登记、备案、许可、监管信息。

名称 佛山市康复医疗设备厂

注册资金 人民币伍拾万元

类型 集体所有制

成立日期 1995年05月30日

法定代表人 黄伟民

经营期限 长期

经营范围 生产、销售：Ⅲ类6864医用卫生材料及敷料，Ⅱ类6866医用高分子材料及制品（凭有效许可证经营）。制造、加工、修理：轮椅、拐杖、残疾人机动车；加工：塑料，纺织品；货物或技术进出口（国家禁止或涉及行政审批的货物和技术进出口除外）。加工：塑料，纺织品。☐（依法须经批准的项目，经相关部门批准后方可开展经营活动。）☐

住所 佛山市高明区荷城街道富湾高富一路



登记机关

2020



2020年3月18日

http://www.gsxt.gov.cn

国家企业信用信息公示系统报送公示年度报告

国家市场监督管理总局监制

国家企业信用信息公示系统网址：

# 医疗器械生产许可证

许可证编号：粤食药监械生产许20010139号


企业名称：佛山市康复医疗设备厂

生产地址：佛山市高明区荷城街道富湾高富一路

法定代表人：黄伟民

生产范围：见医疗器械生产产品登记表

企业负责人：黄伟民



住所：佛山市高明区荷城街道富湾高富一路

发证部门：广东省食品药品监督管理局





有效期限：至 2020 年 11 月 15 日

发证日期：2016 年 12 月 01 日

国家食品药品监督管理总局制

# 医疗器械生产产品登记表

企业名称	佛山市康复医疗设备厂			
许可证编号	粤食药监械生产许 20010139 号			
许可证有效期限	至 2020 年 11 月 15 日			
生产范围	II 类 6864 医用卫生材料及敷料, II 类 6866 医用高分子材料及制品			
<b>生产产品列表</b>				
序号	产品名称	注册号	登载日期	备注
1	外科纱布敷料	粤械注准 20152640596	2015 年 11 月 16 日	
2	一次性使用医用口罩	粤械注准 20152641443	2016 年 06 月 06 日	
3	一次性使用手术衣	粤械注准 20172641690	2018 年 08 月 08 日	
4	一次性使用无菌导尿管	粤械注准 20172661795	2018 年 08 月 08 日	
5	医用外科口罩	粤械注准 20172641484	2018 年 08 月 08 日	
发证部门(公章):				



佛山市康复医疗设备厂


# 中华人民共和国医疗器械注册证

注册证编号：粤械注准 20172641484

注册人名称	佛山市康复医疗设备厂
注册人住所	佛山市高明区荷城街道富湾高富一路
生产地址	佛山市高明区荷城街道富湾高富一路
产品名称	医用外科口罩
型号、规格	型号：挂耳式、绑带式；规格： $(18 \pm 1)$ cm $\times$ $(9.5 \pm 0.5)$ cm
结构及组成	由罩体、口罩带、鼻夹组成。
适用范围	适用于医务人员或相关人员的基本防护，以及在有创操作过程中阻止血液、体液和飞溅物传播的防护。
附件	产品技术要求。
其他内容	无
备注	原产品注册证号：粤食药监械（准）字 2014 第 264031 号

审批部门：广东省食品药品监督管理局

批准日期：2017年08月22日  
有效期至：2022年08月22日



# 对外贸易经营者备案登记表

统一社会信用代码: 914406081935502519

备案登记表编号: 02477932

进出口企业代码: \_\_\_\_\_

经营者中文名称	佛山市康复医疗设备厂		
经营者英文名称	Foshan Kangfu Medical- Facility Factory		
组织机构代码	_____	经营者类型 (由备案登记机关填写)	集体企业
住所	广东省佛山市高明区荷城街道富湾镇高富一路		
经营场所(中文)	广东省佛山市高明区荷城街道富湾镇高富一路		
经营场所(英文)	Gaofu 1st Road, Fuwan Town, Hecheng Street, Gaoming District, Foshan City, Guangdong Province		
联系电话	0757-88818892	联系传真	0757-88819028
邮政编码	528500	电子邮箱	347592356@qq.com
工商登记注册日期	1995-5-30	工商登记注册号	_____

依法办理工商登记的企业还须填写以下内容


企业法定代表人姓名	黄伟民	有效证件号	440601196510050613
注册资金	伍拾万元	(折美元)	

依法办理工商登记的外国(地区)企业或个体工商户(独资经营者)还须填写以下内容

企业法定代表人/ 个体工商户负责人姓名	_____	有效证件号	_____
企业资产/个人财产	_____	(折美元)	

备注	_____
----	-------

填表前请认真阅读背面的条款,并由企业法定代表人或个体工商户负责人签字、盖章。



2020



## 本对外贸易经营者作如下保证：

- 一、遵守《中华人民共和国对外贸易法》及其配套法规、规章。
  - 二、遵守与进出口贸易相关的海关、外汇、税务、检验检疫、环保、知识产权等中华人民共和国其他法律、法规、规章。
  - 三、遵守中华人民共和国关于核、生物、化学、导弹等各类敏感物项和技术出口管制法规以及其他相关法律、法规、规章，不从事任何危害国家和社会公共利益的活动。
  - 四、不伪造、变造、涂改、出租、出借、转让、出卖《对外贸易经营者备案登记表》。
  - 五、在备案登记表中所填写的信息是完整的、准确的、真实的；所提交的所有材料是完整的、准确的、合法的。
  - 六、《对外贸易经营者备案登记表》上填写的任何事项发生变化之日起，30日内到原备案登记机关办理《对外贸易经营者备案登记表》的变更手续。
- 以上如有违反，将承担一切法律责任。

对外贸易经营者签字 盖章

2020年3月20日

**注：**1、备案登记表中“组织机构代码”一栏，由企业、组织和取得组织机构代码的个体工商户填写。

2、依法办理工商登记的外国（地区）企业，在经营活动中，承担有限 / 无限责任。  
依法办理工商登记的个体工商户（独资经营者），在经营活动中，承担无限责任。

3、工商登记营业执照中，如经营范围不包括进口商品的分销业务，备案登记机关应在备注栏中注明“无进口商品分销业务”。

## 海关进出口货物收发货人备案回执

企业名称	佛山市康复医疗设备厂
统一社会信用代码	914406081935502519
海关编码	440695003A
检验检疫备案号	4456200538
有效期	长期

中华人民共和国  
(注册海关) 高明海关  
(注册日期) 2020年04月13日  
注册备案专用章

自然人、法人或者非法人组织可通过“中国海关企业进出口信用信息公示平台” (<http://credit.customs.gov.cn>) 或者“互联网+关” (<http://online.customs.gov.cn>) 查询海关公示的企业信息。





European Commission

# Declaration of Conformity

**Manufacturer:** Foshan Kangfu Medical- Facility Factory

**Address:** Gaofu 1st Road, Fuwan Town, Hecheng Street, Gaoming District, Foshan City, Guangdong Province, 528531, China

**EU Authorised Representative:**

Osmunda Medical Technology Service GmbH

Address: Von Oppen-Weg 15, 14476 Potsdam, Germany

DIMDI code: DE/0000047267

**Device:** Surgical medical mask    **Model:** Ear hanging type

**Classification (MDR, Annex VIII):** Class I

**Conformity assessment route:** ANNEX II+ANNEX III

We herewith declare that the above mentioned product meet the provisions of the following regulation (EU) MDR 2017/745. All supporting documentations are retained under the premises of the manufacturer.



**General applicable regulation:**

REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on medical devices.



**Standard Applied:**

EN ISO 13485:2016	EN ISO 14971:2012	EN 14683:2019 +AC: 2019	BS EN ISO 15223-1:2016
EN ISO 10993-1:2009/AC:2010	EN ISO 10993-5:2009	ISO 10993-10:2010	EN 1041:2008

**Place, Date of Issue:**

Foshan, 2<sup>nd</sup> April 2020

Signature:

Management Representative



# Declaration of Conformity

**Manufacturer:** Foshan Kangfu Medical- Facility Factory

**Address:** Gaofu 1st Road, Fuwan Town, Hecheng Street, Gaoming District, Foshan City, Guangdong Province, 528531, China

**EU Authorised Representative:**

Osmunda Medical Technology Service GmbH

Address: Von Oppen-Weg 15, 14476 Potsdam, Germany

DIMDI code: DE/0000047267

**Device:** Surgical medical mask      **Model:** Ear hanging type

**Classification (MDR, Annex VIII):** Class I

**Conformity assessment route:** ANNEX II+ANNEX III

We herewith declare that the above mentioned product meet the provisions of the following regulation (EU) MDR 2017/745. All supporting documentations are retained under the premises of the manufacturer.



**General applicable regulation:**

REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on medical devices.



**Standard Applied:**

EN ISO 13485:2016      EN ISO 14971:2012      EN 14683:2019      BS EN ISO 15223-1:2016

+AC: 2019

EN ISO 10993-1:2009/AC:2010      EN ISO 10993-5:2009      ISO 10993-10:2010      EN 1041:2008

**Place, Date of Issue:**

Foshan, 2nd April, 2020

**Signature: Management Representative**





**Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG**  
**General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG**

**Formblatt für Medizinprodukte, außer In-vitro-Diagnostika**  
**Form for Medical Devices except In Vitro Diagnostic Medical Devices**

<b>Zuständige Behörde / Competent authority</b>	
Code DE/CA76	
Bezeichnung / Name Landesamt für Arbeitsschutz, Verbraucherschutz und Gesundheit, Abteilung Gesundheit, Dezernat G4	
Staat / State Deutschland	Land / Federal state Brandenburg
Ort / City Zossen	Postleitzahl / Postal code 15806
Straße, Haus-Nr. / Street, house no. Wünsdorfer Platz 3	
Telefon / Phone +49-331-8683852	Telefax / Fax +49-331-8683865
E-Mail / E-mail medizinprodukte@lavg.brandenburg.de	

OSM  
Osmunda Me

<b>Anzeige / Notification</b>	
Registriertdatum bei der zuständigen Behörde Registration date at competent authority	Registriernummer / Registration number
Typ der Anzeige / Notification type <input checked="" type="checkbox"/> Erstanzeige / Initial notification <input type="checkbox"/> Änderungsanzeige / Notification of change <input type="checkbox"/> Widerrufsanzeige / Notification of withdrawal	
Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn	
Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG <input type="checkbox"/> Hersteller / Manufacturer <input checked="" type="checkbox"/> Bevollmächtigter / Authorised Representative <input type="checkbox"/> Einführer / Importer <input type="checkbox"/> Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG <input type="checkbox"/> Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV <input type="checkbox"/> Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG	





Anzeigender / Reporting organisation (person)	
Code	DE/0000047267
Bezeichnung / Name	Osmunda Medical Technology Service GmbH
Staat / State	Deutschland
Land / Federal state	Brandenburg
Ort / City	Potsdam
Postleitzahl / Postal code	14476
Straße, Haus-Nr. / Street, house no. Von Oppen-Weg, 15	
Telefon / Phone	030 50590627
Telefax / Fax	
E-Mail / E-mail	eu@osmundacn.com

**iUNDA**  
 dical Tech  
 Handeire


Hersteller / Manufacturer	
Bezeichnung / Name	Foshan Kangfu Medical- Facility Factory
Staat / State	CN
Ort / City	Foshan, Guangdong
Postleitzahl / Postal code	528531
Straße, Haus-Nr. / Street, house no. Gaofu 1st Road, Fuwan Town, Hecheng Street, Gaoming District	
Telefon / Phone	0086 -0757-86430687
Telefax / Fax	
E-Mail / E-mail	

Sicherheitsbeauftragter für Medizinprodukte nach § 30 Abs. 2 MPG 9) Safety officer for medical devices pursuant to § 30 (2) Medical Devices Act, MPG	
Bezeichnung / Name	Min Yang
Staat / State	Deutschland
Land / Federal state	Berlin
Ort / City	Berlin
Postleitzahl / Postal code	10787
Straße, Haus-Nr. / Street, house no. Keithstr. 2-4	
Telefon / Phone	030 5059 0627
Telefax / Fax	
E-Mail / E-mail	min.yang@osmundacn.com




<b>Vertreter / Deputy (optional)</b>		
Bezeichnung / Name		
Telefon / Phone		Telefax / Fax
E-Mail / E-mail		
<input type="checkbox"/> Erstanzeige / Initial notification <input checked="" type="checkbox"/> Änderungsanzeige / Notification of change		

佛山供药者 不作其他用途



Medizinprodukt (Erstmaliges Inverkehrbringen) / Medical device (First placing on the market)	
Klasse / Class	<input checked="" type="checkbox"/> I <input type="checkbox"/> I - steril / sterile <input type="checkbox"/> I - mit Messfunktion / with measuring function <input type="checkbox"/> I - steril und mit Messfunktion / sterile and with measuring function <input type="checkbox"/> IIa <input type="checkbox"/> IIb <input type="checkbox"/> III <input type="checkbox"/> III - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012 <input type="checkbox"/> Aktives implantierbares Medizinprodukt / Active implantable medical device <input type="checkbox"/> Aktives implantierbares Medizinprodukt - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012 Active implantable medical device - manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012
App (Software auf mobilen Endgeräten)	<input type="checkbox"/> ja / yes <input checked="" type="checkbox"/> nein / no
Nummer(n) der Bescheinigung(en) / Certificate number(s)	
Handelsname des Produktes / Trade name of the device	Surgical medical mask
Produktbezeichnung / Name of device	
Nomenklaturcode / Nomenclature code	12-458
Nomenklaturbezeichnung / Nomenclature term	Maske, Chirurgie
Kategoriecode / Category code	10
Kategorie / Category	Produkte zum Einmalgebrauch
Kurzbeschreibung deutsch / German short description	
Kurzbeschreibung englisch / English short description	The surgical medical mask is applicable to the basic protection of medical personnel or related personnel, as well as the protection against the spread of blood, body fluids and splashes in the process of invasive operation.

nology S  
 gister-Nr.:  
 www.bs



Medizinprodukte (Aufbereiten) / Medical devices (Reprocessing)	
<input type="checkbox"/>	Semikritische Medizinprodukte / Semicritical medical devices <input type="checkbox"/> Gruppe A / Group A <input type="checkbox"/> Gruppe B / Group B
<input type="checkbox"/>	Kritische Medizinprodukte / Critical medical devices <input type="checkbox"/> Gruppe A / Group A <input type="checkbox"/> Gruppe B / Group B <input type="checkbox"/> Gruppe C / Group C Nummer der Bescheinigung / Certificate number
<input type="checkbox"/>	Sterilisationsverfahren / Sterilisation procedures <input type="checkbox"/> Dampfsterilisation / Steam sterilisation <input type="checkbox"/> Gassterilisation / Gas sterilisation <input type="checkbox"/> Strahlensterilisation / Radiation sterilisation <input type="checkbox"/> andere / others Angewandtes Verfahren / Applied procedure

Service GmbH  
 HRB 28353 P  
 osmundacn.com

Ich versichere, dass die Angaben nach bestem Wissen und Gewissen gemacht wurden.  
 I affirm that the information given above is correct to the best of my knowledge.

Ort  
 City

Berlin

Datum  
 Date

Name



2020-04-05



Osmunda Medical Technology Service GmbH

Handelregister-Nr.: HRB 28353 P

Min Yang w.osmundacn.com

Unterschrift  
 Signature

Bearbeitungsvermerke / Processing notes Nur von der zuständigen Behörde auszufüllen / To be filled in only by the competent authority	
Bearbeiter / Person responsible	Telefon / Phone




# Foshan Kangfu Medical- Facility Factory

## COA

No: 20200418-1

Test Number: 20200412

Sample Name:	Surgical medical mask	Number of sample:	60
Production unit:	Foshan Kangfu Medical- Facility Factory	Sample Character:	Solid
Trademark:	 KangFu KangFu	Disinfection date:	2020/4/12
Inspection date:	2020-4-13	Sample Receiving Date:	2020-4-13
Inspection department:	laboratory		

Test results:

Test items	Test Results	Reference Standards
Bacterial filtration efficiency (BFE), %	98.30%	≥ 98
Differential pressure (Pa/cm <sup>2</sup> )	48.2	< 60
Splash resistance pressure (kPa)	16	≥ 16.0
Microbial cleanliness (cfu/g)	20	≤ 30



Examiner: Jinghua Guan      Inspector: Jun Yang      Leader: XianZhi Huang  
 April 18, 2020      April 18, 2020      April 18, 2020

Remarks:

- The COA will be invalid when it is altered, added and deleted, partially copied and not stamped the COA seal.
- The COA shall not be used for commercial advertising.
- The COA is in duplicate, one for the supplier, the other is deposited in the laboratory.

康孚医疗



# Certificate of conformity

Product name: Surgical Medical Mask

Model: Ear hanging type

Specifications: (18±1)cm × (9.5±0.5)cm

Quantity: 10 pcs/pack

Production batch number: 20200412

Production date: April 17, 2020

Expiration date: April 1, 2022

Inspector: Inspection

**qualified**

Executive standard: EN 14683-2019+AC: 2019

**Foshan Kangfu Medical Facility Factory**



佛山康复医疗设施有限公司

佛山康复医疗设施有限公司



KangDeFu Kindful

# Disposable Protective Mask

Three layers with breathable comfort



50 pcs/box



Breathe  
Comfortably



Professional  
Protection



Safe and  
Healthy




Three  
Layers



LOWER AIR RESISTANCE, EASIER BREATHING

# Powerful Barriers Easy and Free Breathing

Filtering out 99.99% of bacteria, dust and particles




## **WATERPROOF NON-WOVEN FABRIC**

BLOCKING INFILTRATION  
OF MIST AND WATER  
VAPOR



## **MEDICAL GRADE MELTBLOWN FABRIC**

BLOCKING FINE DUST  
AND PARTICLES




## **SKIN-FRIENDLY NON-WOVEN FABRIC**

SKIN FRIENDLY  
COMFORTABLE SWEAT  
AND OIL-ABSORBING



KangDeFu Kindful

# Three-layer High-tech Protection




KangDeFu Kindful

# Fitting facial contours

## Top Face fitting system

Three-dimensional cutting, Three-dimensional shape, more comfortable to wear



### Face Fitting Design

Popular mask design, suitable for most people

### Memory Nose Wire

Nose wires are made of memory metal to effectively fix the mask

### Comfortable Ear-loops

Uniquely customized ear-loops prevent any harms to your ears

# Product Specs

<b>PRODUCT:</b>	WENXIANG I-CLEAR
<b>FEATURES:</b>	DISPOSABLE PROTECTIVE MASK (3 LAYERS)
<b>STANDARD</b>	Q/XHB001-2020
<b>PROTECTION:</b>	LEVEL A
<b>QUANTITY</b>	50PCS / BOX
<b>SPECIFICATIONS</b>	18.0 X 9.5 cm 3 LAYERS



KangDeFu Kindful







MEAS : 645\*490\*400MM



**营 业 执 照**

(副 本)(副本号:2-1)

统一社会信用代码  
914406081935502519

扫描二维码登录“  
国家企业信用信息公示系统”了解更  
多登记、备案、许  
可、监管信息。

名 称 佛山市康复医疗设备厂      注 册 资 金 人民币伍拾万元

类 型 集体所有制      成 立 日 期 1995年05月30日

法 定 代 表 人 黄伟民      经 营 期 限 长期

经 营 范 围 生产、销售：Ⅲ类6864医用卫生材料及敷料，Ⅱ类6866  
6医用高分子材料及制品（凭有效许可证经营）。制造  
、加工、修理：轮椅、拐杖、残疾人机动车；加工：  
塑料，纺织品；货物或技术进出口（国家禁止或涉及  
行政审批的货物和技术进出口除外）。加工：塑料，  
纺织品。■（依法须经批准的项目，经相关部门批准  
后方可开展经营活动。）■

住 所 佛山市高明区荷城街道富湾高富  
一路

登 记 机 关 

2020 年 1 月 18 日

http://www.gsxt.gov.cn  
国家企业信用信息公示系统网址：  
非市场主体请于每年1月1日至3月31日期间  
国家企业信用信息公示系统报送公示年度报告  
国家市场监督管理总局监制

**医 疗 器 械 生 产 许 可 证**

许可证编号：粤食药监械生产许20010139号

企业名称：佛山市康复医疗设备厂      生产地址：佛山市高明区荷城街道富湾高富一路

法定代表人：黄伟民      生产范围：见医疗器械生产产品登记表

企业负责人：黄伟民

住 所：佛山市高明区荷城街道富湾高富一路      发证部门：广东省食品药品监督管理局

有效期限：至 2020 年 11 月 15 日      发证日期：2016 年 12 月 01 日

国家食品药品监督管理局制



## 医疗器械生产产品登记表

<b>企业名称</b>	佛山市康复医疗设备厂			
<b>许可证编号</b>	粤食药监械生产许 20010139 号			
<b>许可证有效期限</b>	至 2020 年 11 月 15 日			
<b>生产范围</b>	Ⅱ类 6864 医用卫生材料及敷料, Ⅱ类 6866 医用高分子材料及制品			
<b>生产产品列表</b>				
<b>序号</b>	<b>产品名称</b>	<b>注册号</b>	<b>登载日期</b>	<b>备注</b>
1	外科纱布敷料	粤械注准 20152640596	2015 年 11 月 16 日	
2	一次性使用医用口罩	粤械注准 20152641443	2016 年 06 月 06 日	
3	一次性使用手术衣	粤械注准 20172641690	2018 年 08 月 08 日	
4	一次性使用无菌导尿包	粤械注准 20172661795	2018 年 08 月 08 日	
5	医用外科口罩	粤械注准 20172641487	2018 年 08 月 08 日	
<b>发证部门(公章):</b> <div style="text-align: center;">  <p>2018 年 08 月 08 日</p> </div>				

## 对外贸易经营者备案登记表

统一社会信用代码: 914406081935502519

备案登记表编号: 02477932

进出口企业代码: \_\_\_\_\_

经营者中文名称	佛山市康复医疗设备厂		
经营者英文名称	Foshan Kangfu Medical- Facility Factory		
组织机构代码	_____	经营者类型 (由备案登记机关填写)	集体企业
住 所	广东省佛山市高明区荷城街道富湾镇高富一路		
经营场所 (中文)	广东省佛山市高明区荷城街道富湾镇高富一路		
经营场所 (英文)	Gaofu 1st Road, Fuwan Town, Hecheng Street, Gaoming District, Foshan City, Guangdong Province		
联系电话	0757-88818892	联系传真	0757-88819028
邮政编码	528500	电子邮箱	347592356@qq.com
工商登记注册日期	1995-5-30	工商登记注册号	_____

依法办理工商登记的企业还须填写以下内容

企业法定代表人姓名	黄伟民	有效证件号	440601196510050613
注册资金	伍拾万元		(折美元)

依法办理工商登记的外国 (地区) 企业或个体工商户 (独资经营者) 还须填写以下内容

企业法定代表人/ 个体工商户负责人姓名	_____	有效证件号	_____
企业资产/个人财产	_____		(折美元)

备注	_____
----	-------

填表前请认真阅读背面的条款, 并由企业法定代表人或个体工商户负责人签字、盖章。





2020



# 中华人民共和国医疗器械注册证

注册证编号：粤械注准 20172641484

注册人名称	佛山市康复医疗设备厂
注册人住所	佛山市高明区荷城街道富湾高富一路
生产地址	佛山市高明区荷城街道富湾高富一路
产品名称	医用外科口罩
型号、规格	型号：挂耳式、绑带式；规格：(18±1) cm × (9.5±0.5) cm
结构及组成	由罩体、口罩带、鼻夹组成。 
适用范围	适用于医务人员或相关人员的基本防护，以及在有创操作过程中阻止血液、体液和飞溅物传播的防护。
附件	产品技术要求。
其他内容	无
备注	原产品注册证号：粤食药监械（准）字 2014 第 264031A 号 

审批部门：广东省食品药品监督管理局

批准日期：2017年08月21日

有效期至：2022年08月21日



European  
Commission

## Declaration of Conformity

**Manufacturer:** Foshan Kangfu Medical- Facility Factory

**Address:** Gaofu 1st Road, Fuwan Town, Hecheng Street, Gaoming District, Foshan City, Guangdong Province, 528531, China

**EU Authorised Representative:**

Osmunda Medical Technology Service GmbH

Address: Von Oppen-Weg 15, 14476 Potsdam, Germany

DIMDI code: DE/0000047267

**Device:** Surgical medical mask    **Model:** Ear hanging type

**Classification (MDR, Annex VIII):** Class I

**Conformity assessment route:** ANNEX II+ANNEX III

We herewith declare that the above mentioned product meet the provisions of the following regulation (EU) MDR 2017/745. All supporting documentations are retained under the premises of the manufacturer.



**General applicable regulation:**

REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on medical devices.

**Standard Applied:**

EN ISO 13485:2016	EN ISO 14971:2012	EN 14683:2019 +AC: 2019	BS EN ISO 15223-1:2016
EN ISO 10993-1:2009/AC:2010	EN ISO 10993-5:2009	ISO 10993-10:2010	EN 1041:2008

**Place, Date of Issue:**

Foshan, 2<sup>nd</sup> April 2020

Signature:

Management Representative



# HOW TO PROPERLY WEAR



①

Open the mask and keep the skin dry with the white side facing inward and nose wire uppermost.



②

Position the ear loops around both ears and even the stress.



③

Open the mask and keep the skin dry with the white side facing inward and nose wire uppermost.

## Precaution

1. Do not use if the product is presenting bad odor, bleached color or mold.
- 2 This product is not suitable for protection against toxic gases and severe pollution.
- 3 Any use discomfort, requires the to replace his mask with a new one.. This mask is of one-time use. Cleaning is not enough a measure to prevent the spreading of diseases.

## Storage

It needs to be stored in a dry and clean place. It is prohibited by law to mix and store with toxic, harmful and polluting materials



KangDeFu Kindful



**SUBJECT** Physical & Microbiological Test

**TEST LOCATION** TÜV SÜD China

TÜV SÜD Products Testing (Shanghai) Co., Ltd.  
B-3/4, No.1999 Du Hui Road, Minhang District  
Shanghai 201108, P.R. China

**CLIENT NAME** Foshan Kangfu Medical- Facility Factory

**CLIENT ADDRESS** He Cheng Jie Dao Fu Wan Zhen Gao Fu Yi Lu, Gaoming District, Foshan City,  
Guangdong Province, China

**TEST PERIOD** 11-Apr-2020~19-Apr-2020

Prepared By

*Bella Xu*

(Bella Xu)  
Report Drafter

Authorized By



(Leo Liu)  
Authorized Signatory



**Note:** (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested. (3) The test report shall not be reproduced except in full without the written approval of the laboratory. (4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

Chemical/Microbiology Laboratory:  
TÜV SÜD Products Testing (Shanghai) Co.,  
Ltd.  
B-3/4, No.1999 Du Hui Road, Minhang District  
Shanghai  
201108  
P.R. China

Phone : +86 (21) 6037 6375  
Fax : +86 (21) 6037 6345  
Email: food.chem@tuv-sud.cn  
Webpage: www.tuv-sud.cn

Regional Head Office:  
TÜV SÜD Certification and Testing  
(China) Co., Ltd.  
No.151 Heng Tong Road Shanghai  
200 070 P.R.China





### TEST REPORT

Sample Description : surgical face mask  
Sample Quantity : 60 pieces  
Lot Number/Batch Code : 200401  
Specification : Hanging ear type  
Size : /  
Type of Mask : Type IIR  
Brand Name : /


Remark: The above information was provided by applicant.

#### Summary of Test Results

No.	Test Item	Test Standard	Judgement
1	Bacterial Filtration Efficiency (BFE) Test	EN 14683:2019+AC:2019(E) Annex B	Pass
2	Differential Pressure Test	EN 14683:2019+AC:2019(E) Annex C	Pass
3	Synthetic Blood Penetration Test	ISO 22609:2004	Pass
4	Microbial Cleanliness Test	EN 14683:2019+AC:2019(E) Annex D	Pass

Note: Pass = Meet customer requirements;  
Fail = Fail customer requirements;  
# = No comment;  
N.D. = Not detected.

#### Photo of Samples





Results

No.	Test Item	Test Result
1	Bacterial Filtration Efficiency (BFE) Test	Specimen 1#: 99.5% Specimen 2#: 99.5% Specimen 3#: 99.5% Specimen 4#: 99.5% Specimen 5#: 99.6%
2	Differential Pressure Test	42.9 Pa/cm <sup>2</sup>
3	Synthetic Blood Penetration Test	Specimen 1#~13#: None seen
4	Microbial Cleanliness Test	Specimen 1#: <1 CFU/g Specimen 2#: <1 CFU/g Specimen 3#: <1 CFU/g Specimen 4#: <1 CFU/g Specimen 5#: <1 CFU/g

Bacterial Filtration Efficiency (BFE) Test

1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of mask.

2. Sample description was given by client

Sample description : surgical face mask  
Specification : Hanging ear type  
Lot Number : 200401  
Sample Receiving Date : 2020-04-11

3. Test Method

EN 14683:2019+AC:2019(E) Annex B

4. Apparatus and materials

- 4.1 *Staphylococcus aureus* ATCC 6538.
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.

5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.







## 6. Procedure

- 6.1 Preparation of the bacterial challenge: Dilute the culture in peptone water to achieve a concentration of approximately  $5 \times 10^5$  CFU/mL.
- 6.2 Adjust the flow rate through the Anderson sampler to 28.3 L/min.
- 6.3 Deliver the challenge to the nebulizer using a syringe pump. Purge tubing and nebulizer of air bubbles.
- 6.4 Perform a positive control run without a test specimen to determine the number of viable aerosol particles being generated. The mean particle size (MPS) of the aerosol will also be calculated from the results of these positive control plates.
  - 6.4.1 Initiate the aerosol challenge by turning on the air pressure and pump connected to the nebulizer. Immediately begin sampling the aerosol using the Anderson sampler.
  - 6.4.2 Time the challenge suspension to be delivered to the nebulizer for 1 min.
  - 6.4.3 Time the air pressure and Anderson sampler to run for 2 min.
  - 6.4.4 At the conclusion of the positive control run, remove plates from the Anderson sampler.
- 6.5 Place new agar plates into Anderson sampler and clamp the test specimen into the top of the Anderson sampler, with the inside of the specimen facing towards the bacterial challenge (test area:  $77\text{cm}^2$ ).
- 6.6 Repeat the challenge procedure for each test specimen.
- 6.7 Repeat a positive control after completion of the sample set.
- 6.8 Perform a negative control run by collecting a 2 min sample of air from the aerosol chamber. No bacterial challenge should be pumped into the nebulizer during the collection of the negative control.
- 6.9 Incubate agar plates at  $(37 \pm 2)^\circ\text{C}$  for (20 to 52) h.
- 6.10 Count each of the six-stage plates of the Anderson sampler.

## 7. Calculation

Total the count from each of the six plates for the test specimens and positive controls, as specified by the manufacture of Anderson sampler. The filtration efficiency percentages are calculated as follows:

$$\text{BFE} = (C - T) / C \times 100$$

$T$  is the total plate count for the test specimen.

$C$  is the mean of the total plate counts for the two positive controls.





8. Test results\*

Stage Number	P Value	Positive Control (A)	Positive Control (B)	Negative Control	Specimen 1#	Specimen 2#	Specimen 3#	Specimen 4#	Specimen 5#
1		36	48	0	0	0	0	0	0
2		60	70	0	0	0	0	0	0
3		183	328	0	0	0	0	0	0
4		508	610	0	1	6	0	0	0
5		883	793	0	6	5	11	5	8
6		606	703	0	5	1	2	7	2
Total (T), CFU		2276	2552	<1	12	1	13	12	10
Average (C), CFU	$2.4 \times 10^3 = (P_A + P_B) / 2$								
BFE, %					99.5	99.5	99.5	99.5	99.6
Requirements					≥ 98				
Remarks	<p><i>P</i> is the value of corresponding corrected particle counts as specified by the manufacturer of the cascade impactor.  <i>T</i> is the total of <i>P</i> value for the test specimen.  <i>C</i> is the mean of the total of <i>P</i> value of the two positive controls.</p>								





Differential pressure Test

1.Purpose

The purpose of the test was to measure the differential pressure of masks.

2.Sample description was given by client

Sample description : surgical face mask  
Specification : Hanging ear type  
Lot Number : 200401  
Sample Receiving Date : 2020-04-11

3.Test Method

EN 14683:2019+AC:2019(E) Annex C

4. Apparatus and materials

Differential pressure testing instrument

5.Test specimen

- 5.1 Test specimen are complete masks or shall be cut from masks. Each specimen shall be able to provide 5 different circular test areas of 2.5 cm in diameter.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at  $(21 \pm 5)^\circ\text{C}$  and  $(85 \pm 5)\%$  relative humidity.

6. Procedure

- 6.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.
- 6.2 The pretreated specimen is placed across the orifice (total area  $4.9\text{cm}^2$ , test area diameter 25mm) and clamped into place so as to minimize air leaks.
- 6.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.
- 6.4 The differential pressure is read directly.
- 6.5 The procedure described in steps 6.1-6.4 is carried out on 5 different areas of the mask and readings averaged.

Results:

Specimen	Test Results* (Pa/cm <sup>2</sup> )	Average (Pa/cm <sup>2</sup> )	Requirements	Judgement
1#	48.0	42.9	< 60	Pass
2#	46.3			
3#	41.8			
4#	39.4			
5#	39.1			

TIA  
1  
\*

## Synthetic Blood Penetration Test

### 1. Purpose

For evaluation of resistance of masks to penetration by a fixed volume of synthetic blood at a high velocity.

### 2. Sample description was given by client

Sample description : surgical face mask  
Specification : Hanging ear type  
Lot Number : 200401  
Sample Receiving Date : 2020-04-11

### 3. Test Method

ISO 22609:2004

### 4. Apparatus and materials

- 4.1 Synthetic blood.
- 4.2 Tensiometer.
- 4.3 Synthetic blood penetration test apparatus;
- 4.4 Targeting plate.
- 4.5 Air pressure source.
- 4.6 Ruler.
- 4.7 Balance.
- 4.8 Controlled temperature and humidity chamber.

### 5. Test specimen

- 5.1 As requested by client, take a total of 13 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4h at  $(21\pm 5)^{\circ}\text{C}$  and  $(85\pm 5)\%$  relative humidity.

### 6. Procedure

- 6.1 Prepare the synthetic blood (40~44 mN/m) for the test.
- 6.2 Determine the density of the synthetic blood.
- 6.3 Fill the reservoir with new synthetic blood.
- 6.4 Position the test specimen 30.5 cm (12 in.) from the exit of the canula.
- 6.5 Set the reservoir pressure to the approximate pressure.
- 6.6 Place the targeting plate approximately 1 cm away from the mask.
- 6.7 Set the valve timer to 0.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).



6.8 Set the valve timer to 1.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).

6.9 Calculate the difference in weight of the two spurts. For a test fluid with a density of 1.003, Table 1 gives the target difference in weight plus lower and upper limits for a velocity range within 2% of the target.

Table 1 Target weight difference

Fluid Pressure (mmHg)	Weight difference for 1s difference in spurt duration (g)		
	Min.	Target	Max.
120	3.002	3.063	3.124

6.10 Adjust the reservoir pressure and repeat steps 6.7 to 6.9 until the weight difference is within the target range.

6.11 Record the weight difference for the spurts exiting the nozzle.

6.12 Record the pressure in the reservoir.

6.13 Set the valve time to 0.5 s. Collect and weigh the amount of fluid passing through the targeting hole.

6.14 Set the valve time to 1.5 s. Collect and weigh the amount of fluid passing through the targeting hole.

6.15 The difference in weight between the 0.5 s and 1.5 s spurts through the targeting plate shall be within +2 % ~ -5 % of the difference in weight from the nozzle.

6.16 If the differential weight is less than 95 % of the weight difference exiting the nozzle, check the aim of the stream to make sure it is passing cleanly through the targeting hole.

6.17 If the differential weight is more than 102 % of the weight difference exiting the nozzle, repeat the weight measurements exiting the nozzle (steps 6.7 to 6.11).

6.18 For standard synthetic blood, the timer duration can be estimated using the formula:

$$(p \text{ is the density of the test fluid.}) t = 0.5 + (2 \times p - g \text{ at } 0.5 \text{ s}) / (g \text{ at } 1.5 \text{ s} - g \text{ at } 0.5 \text{ s}).$$

6.19 Record the timer setting to use as the starting point for subsequent testing.

6.20 Mount a test specimen on the specimen holding fixture. If the mask contains pleats, spread the pleats out when mounting the mask onto the fixture to present a single layer of material as the target area.

6.21 Squirt the synthetic blood onto the test specimen for the calculated time. Ensure that the synthetic blood hits the target area of mask.

6.22 Inspect the inside surface for synthetic blood penetration within 10 s of squirting the synthetic blood against the target area.

6.23 Report the results (none / penetration) for each test specimen at the test pressure.



**Results:**

Specimen	Test Results*	Requirements	Judgement
1#	None Seen	Pass Pressure at 16.0 kPa (120mmHg)	Pass
2#	None Seen		Pass
3#	None Seen		Pass
4#	None Seen		Pass
5#	None Seen		Pass
6#	None Seen		Pass
7#	None Seen		Pass
8#	None Seen		Pass
9#	None Seen		Pass
10#	None Seen		Pass
11#	None Seen		Pass
12#	None Seen		Pass
13#	None Seen		Pass





## Microbial Cleanliness Test

### 1. Purpose

The purpose of the test was to measure microbial cleanliness of mask.

### 2. Sample description was given by client

Sample description : surgical face mask  
Specification : Hanging ear type  
Lot Number : 200401  
Sample Receiving Date : 2020-04-11

### 3. Test Method

According to EN ISO 11737-1:2018 to determine the microbial cleanliness of mask material, and refer to the procedure as described in EN 14683:2019+AC:2019(E) Annex D

### 4. Apparatus and materials

- 4.1 Orbital shaker.
- 4.2 0.45 um filter.
- 4.3 Tryptic Soy Agar (TSA).
- 4.4 Sabouraud Dextrose Ager (SDA) with chloramphenicol.
- 4.5 Formula of Extraction Liquid: 1g/L peptone, 5g/L NaCl and 2g/L Tween 20.
- 4.6 Extraction apparatus.

### 5. Test specimen

- 5.1 As requested by client, take a total of 5 mask samples.
- 5.2 Mask samples for testing are provided in the original primary packaging.
- 5.3 Condition at (18 to 26)°C and (45 to 65)% relative humidity during testing.

### 6. Procedure

- 6.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.
- 6.2 The mask is aseptically removed from the packaging and placed in a sterile 500 mL bottle containing 300 mL of extraction liquid.
- 6.3 The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm.
- 6.4 After extracting, 100mL of the extraction liquid is filtered through a 0.45 um filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungi enumeration.
- 6.5 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.
- 6.6 Calculate the colonies of each agar plate.

### 7. Calculation

For each test specimen calculate the microbial cleanliness as follows by counting the total colonies of the TSA and SDA plates.



Results\*:

Specimen	Colonies of the TSA Plate	Colonies of the SDA Plate	Microbial Cleanliness, (CFU/g)	Requirements	Judgement
1#	0	0	<1	According to EN ISO 11737-1:2018 the microbial cleanliness of the mask shall be ≤30 CFU/g tested.	Pass
2#	0	0	<1		
3#	0	0	<1		
4#	0	0	<1		
5#	0	0	<1		

Note:

- 1.\*denotes this test was carried out by external laboratory assessed as competent.
- 2.This report is for internal use only such as internal scientific research, education, quality control, product R&D.

-END OF THE TEST REPORT-





